

Instructions for Use

StatSpin VT Centrifuge Model Number V901

This manual is intended for

For Veterinary Use Only

SSVT StatSpin VT for 100 to 240 VAC, 50/60 Hz

SSVT-1 StatSpin VT for 100 to 240 VAC, 50/60 Hz (Included: RT12 and RH12 rotors)

Printed in U.S.A.



Instructions for Use**StatSpin VT Centrifuge Model Number V901**

PN 55-003786-001EE (August 2025)

© 2025 Beckman Coulter, Inc.

All Rights Reserved.

Trademarks

Beckman Coulter, the stylized logo, and the Beckman Coulter product and service marks mentioned herein are trademarks or registered trademarks of Beckman Coulter, Inc. in the United States and other countries.

All other trademarks are the property of their respective owners.

May be covered by one or more patents – see www.beckmancoulter.com/patents.

Find us on the World Wide Web at:

www.beckmancoulter.com

Original Instructions

Revision History 55-003786-001

EE (August, 2025)

- Removed Limited Warranty

ED (December, 2021)

- Deleted: Lipemia Clearing with LipoClear
- Updated the following:
 - Copyright page
 - Symbols and Definitions
 - Safety Notices Warning and Cautions throughout
 - Centrifuge Intended for Use With
 - Instructions for Use
 - Accessories
 - Determination of Packed Red Cell Volume or Microhematocrit
 - Illustrations
 - Installing the Cover
 - Installing the Rotor
 - Limited Warranty and Disclaimer
 - Opening and Closing the Cover
 - Preparation of Urine Sediment for Microscopic Examination
 - Procedure for Packed Red Cell Volume (Microhematocrit)
 - Quality Control
 - RH12 Microhematocrit Rotor
 - RT12 Tube Rotor
 - Specifications
 - Specimen Processing

EC (September, 2018)

- Moved: Symbol/Regulatory Mark and a link to the website in the California Proposition 65 statement

EB (March, 2018)

- Added: Symbols and Definitions table
- Updated: Manufacturer address, Warning and Cautions, and Limited Warranty statement
- Deleted: CE, and EC Rep

EA (September, 2016)

- Converted the VT Centrifuge Operator's Manual to a Beckman Coulter Instructions for Use (IFU)
- Made general clarifications to the IFU
- Updated Logo
- Added CE, and EC Rep
- Added warning and caution statements, and updated existing warning and caution statements
- Added the Recycling Label information
- Updated specifications, symbol tables, error codes, troubleshooting and maintenance sections
- Updated Limited Warranty statement

Safety Notice

Read all product manuals and consult with Beckman Coulter-trained personnel before you operate the system. Do not perform any procedure before you carefully read all instructions. Always follow the product labels and the manufacturer's recommendations. If you have any questions:

- Visit <http://www.beckmancoulter.com>.
- US customers: Contact Beckman Coulter Customer Support at 1-800-854-3633.
- International customers: Contact your local distributor.

Alerts for Warning, Caution, Important, Note, and Tip

Warning

Warning indicates a potentially hazardous situation which can cause death or serious injury. Warning can indicate the possibility of erroneous data that could cause an incorrect diagnosis.

Caution

Caution indicates a potentially hazardous situation which can cause minor or moderate injury. Caution can also alert against unsafe practices, or indicate the possibility of erroneous data that could cause an incorrect diagnosis.

Important

Important indicates important information to follow.

Note

Note indicates notable information to follow.

Tip

Tip indicates information to consider.

Warnings and Cautions

Pay close attention to the instructions that accompany the notes and symbols and the standard laboratory procedures outlined by your facility and local regulatory agencies.



Warning

Always operate the system with all shields (when used with RT12 rotor) and covers in position and secured to avoid injury.

Perform system operations with caution.

Wear Personal Protective Equipment (PPE) such as gloves, eye shields, and lab coats.

Wash hands thoroughly after contact with sample media and all maintenance activities.

Observe all laboratory policies and procedures related to the handling of biohazardous materials.

Refer to the applicable sources (such as Safety Data Sheets) for specific hazard information.



Warning

Do not expose the rotor to strong or concentrated acids, bases, esters, aromatic or halogenated hydrocarbons, ketones, or strong oxidizing agents, or environmental influences, including natural ultra-violet radiation. Doing so will subject the rotor to corrosion or weakening of the construction.



Warning

Do not operate the centrifuge below the minimum operating temperature. See [Specifications](#).



Warning

Handle and dispose of sharp fragments according to the World Health Organization's Laboratory Biosafety Manual and relevant local and national regulations.



Warning

Safety protection may be impaired if the equipment is used in a manner not specified by the manufacturer.



Warning

Inspect the instrument for cracks or any physical damage to housing, cover, and rotor upon the receipt of the unit. Damage can cause unsafe operation; if damage or cracks are found, discontinue use until repairs have been performed.



Warning

Only reset the cycle counter after conducting the recommended inspections and service. Resetting the cycle counter without performing the recommended inspections and service reduces the reliability and safety of the instrument.



Warning

Only use the external power supply (Beckman Coulter Part Number X01-003553-001) included with the unit. Use of any other external power supply can cause damage to the unit and will void the warranty.



Warning

Outside of North America: do not use the power cord supplied. Use power cord for at least 1.0 Amp with an IEC320/CEE22 female connector and male connector suitable for the power outlet to be used.



Warning

Picking up or moving the centrifuge during operation can cause injury to the operator and/or damage to the centrifuge.



Warning

Electromagnetic Compatibility

This device complies with the emissions and immunity requirements as specified in the EN/IEC 61326 series of Product Family Standards for a "basic electromagnetic environment." Such equipment is supplied directly at low voltage from public mains network. This equipment is not intended for residential use.

This device generates, uses, and can radiate unintentional radio-frequency (RF) energy. If this device is not installed and operated correctly, this RF energy can cause interference with other equipment. It is the responsibility of the end user to be sure that a compatible electromagnetic environment for the device can be maintained so that the device operates as intended.

This equipment is designed for use in a PROFESSIONAL HEALTHCARE FACILITY ENVIRONMENT. It is likely to perform incorrectly if used in a HOME HEALTHCARE

Safety Notice

Warnings and Cautions

ENVIRONMENT. If it is suspected that performance is affected by electromagnetic interference, correct operation may be restored by increasing the distance between the equipment and the source of the interference.

In addition, other equipment can radiate RF energy to which this device is sensitive. If one suspects interference between this device and other equipment, Beckman Coulter recommends the following actions to correct the interference:

- Evaluate the electromagnetic environment before installation and operation of this device.
- Do not operate this device close to sources of strong electromagnetic radiation (for example: unshielded intentional RF sources), as these can interfere with proper operation. Examples of unshielded intentional radiators are handheld radio transmitters, cordless phones, and cellular phones.
- Do not place this device near medical electrical equipment that can be susceptible to malfunctions caused by close-proximity to electromagnetic fields.
- This device has been designed and tested to CISPR 11, Class A emission limits. In a domestic environment, this device can cause radio interference, in which case, you need to take measures to mitigate the interference.

Caution

Disconnect the power cord of the external power supply from the electrical outlet before performing maintenance or inspection.

Caution

Do not leave any rotor on the rotor-holder when the centrifuge is not in use for an extended period of time. Doing so may compress the O-Ring and decrease its ability to hold rotors.

Caution

Do not operate the centrifuge without a shield in place when using an RT12 rotor.

Caution

Do not spray cleaning solutions directly onto the centrifuge bowl or housing. Overspray can reach the motor bearings or internal circuitry, causing harm to the electronics and could also cause corrosion or weakening of the construction of the protective casing.

Before using any cleaning or decontamination methods other than those recommended by the manufacturer, users should check with the manufacturer that the proposed method will not damage the equipment. Cleaning and decontamination

may be necessary as a safeguard before laboratory centrifuges, rotors, and accessories are maintained, repaired, or transferred.

 Caution

Do not use glass tubes of any kind in the RT12 rotor.

 Caution

Do not use the shield with the RH12 Microhematocrit rotor.

 Caution

During operation maintain a 30 cm (12 inch) clearance around the centrifuge. The clearance must be free from obstruction and away from the edge of the surface that the centrifuge is on. The appliance coupler on the external power supply is considered the disconnect device, ensure that it is accessible after installation.

 Caution

If rotor is left in position between runs, be certain to completely seat the bottom of the rotor on the holder before spinning another sample. Failure to correctly seat the rotor each time can cause the rotor to become loose during centrifugation.

 Caution

Follow Universal Precautions with all biological specimens, regardless of whether the specimen is known to contain an infectious agent. (See [References](#))

 Caution

Inspect rotor on a routine basis. Rotor lifespan depends on usage. Inspect the rotor for cracks and replace the rotor immediately when any crack or visible wear occurs.

 Caution

Never operate the centrifuge without the rotor properly mounted. Failure to install and secure the rotor correctly can damage the centrifuge.

 Caution

Position the RT12 rotor in the centrifuge first, then install the shield.

Safety Notice

Warnings and Cautions



Caution

Running the centrifuge repeatedly with an unbalanced load condition can cause excessive vibrations and premature equipment failure.



Caution

The cover interlock bypass is for emergency use only. Disconnect the power cord of the external power supply from the electrical outlet and ensure the rotor has come to a complete stop before using the interlock bypass. If the equipment is not used correctly, safety can be impaired.



Caution

The instructions prohibit use of the specified materials within the centrifuge

- flammable or explosive materials.
- materials which could react chemically with sufficient vigor to cause a hazard.



Caution

The RT12 and RH12 rotors have a finite lifespan that is dependent on usage. The RT12 and the shield should be replaced after approximately 3600 cycles, which is equivalent to 18 months of service running an average of 10 cycles per day. The RH12 rotor should be replaced after approximately 2600 cycles, which is equivalent to 12 months of service running an average of 10 cycles per day. Rotors should be inspected routinely and replaced immediately when any crack or visible wear occurs.



Caution

The centrifuge is designed to use only the RT12 and RH12 rotors. The use of any other rotor may result in a hazard.



Caution

The RT12 rotor must be balanced before operation. If only one sample is being processed, a second sample can serve as the balance tube, or use a similar tube filled with water as a balance.

Please use the instrument as intended. Improper use may cause damage to the instrument, inaccurate results, or potentially nullify warranties.

Symbols and Definitions

Table 1 VT Symbols Glossary

Symbol	Description
	Warning; Biological hazard This symbol indicates a warning of a biological hazard. <i>ISO 7010. Graphical Symbols - Safety colors and safety signs. #W009</i> Supplemental Product-Specific Manufacturer Information This symbol indicates a caution to operate only with all covers in position to decrease risk of personal injury or biohazard. This symbol indicates the use of biohazardous materials in the area. Use caution when working with possible infectious samples. Wear Personal Protective Equipment (PPE) such as gloves, eye shields, and lab coats. Handle and dispose of biohazardous materials according to your laboratory procedures.
	Consult instructions for use This symbol indicates the need for the user to consult the instructions for use. <i>ISO 15223-1. Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General Requirements. #5.4.3</i>
	Caution This symbol indicates the need for the user to consult the instructions for use for important cautionary information such as warnings and precautions that cannot, for a variety of reasons, be presented on the medical device itself. <i>ISO 15223-1. Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General Requirements. #5.4.4</i>

Table 1 VT Symbols Glossary (Continued)

Symbol	Description
	Warning; Crushing of hands This symbol indicates a warning of a closing motion of mechanical parts of equipment. <i>ISO 7010. Graphical Symbols for electrical equipment in medical practices. #W024</i> Supplemental Product-Specific Manufacturer Information Use caution to avoid injury to hands when close to equipment with moving mechanical parts.
	Moving Parts Symbol This symbol indicates that there are moving parts in the area. Only operate the system when all covers are in position and use caution to reduce the risk of personal injury. While the system is operating, do not touch the moving parts of the system. Do not insert fingers or hands into any system opening.
	cNRTLus Certification Mark This symbol indicates recognition by a Nationally Recognized Testing Laboratory (NRTL) that the system has met the relevant product safety standards for the United States and Canada. <i>OSHA, CEC</i>
	RCM Symbol This symbol indicates compliance with the Australian Communications Media Authority (ACMA) requirements (safety and EMC) for Australia and New Zealand.

Table 1 VT Symbols Glossary (Continued)

Symbol	Description
	Recycling Symbol This symbol is required by the Waste Electrical and Electronic Equipment (WEEE) Directive of the European Union. This symbol indicates that: 1. The device was put on the European Market after August 13, 2005. 2. The device is not to be disposed of via the municipal waste collection system of any member state of the European Union. Customers must understand and follow all laws regarding the correct decontamination and safe disposal of electrical equipment. For Beckman Coulter products bearing this label, contact your dealer or your local Beckman Coulter Representative for more information on the take-back program that facilitates the correct collection, treatment, recovery, recycling, and safe disposal of these products. <i>EU Directive 2002-96-EC: waste electrical and electronic equipment (WEEE)</i> For the Japan market: This system is considered an industrial waste, subject to special controls for infectious waste. Before disposal of the system, refer to the <i>Waste Disposal and Public Cleaning Law</i> for compliance procedures.
	RoHS Caution Symbol This symbol indicates that this electronic information product contains certain toxic or hazardous elements, and can be used safely during its environmental protection use period. The number in the middle of the logo indicates the environmental protection use period (in years) for the product. The outer circle indicates that the product can be recycled. The logo also signifies that the product should be recycled immediately after its environmental protection use period has expired. The date on the label indicates the date of manufacture. These labels and materials declaration table (the Table of Hazardous Substance's Name and Concentration) meet People's Republic of China Electronic Industry Standard SJ/T11364-2006 <i>Marking for Control of Pollution Caused by Electronic Information Products</i> requirements.
	"OFF" (power) This symbol indicates disconnection from the mains, at least for mains switches or their positions, and all those cases where safety is involved. <i>IEC 60417: Graphical symbols for use on equipment - Overview and application, #5008</i> Supplemental Product-Specific Manufacturer Information This symbol indicates the off position.

Table 1 VT Symbols Glossary (Continued)

Symbol	Description
	"ON" (power) This symbol indicates connection to the mains, at least for mains switches or their positions, and all those cases where safety is involved. <i>IEC 60417: Graphical symbols for use on equipment - Overview and application, #5007</i> Supplemental Product-Specific Manufacturer Information This symbol indicates the on position.
	Alternating current This symbol indicates on the rating plate that the equipment is suitable for alternating current only; to identify relevant terminals. <i>IEC 60417: Graphical symbols for use on equipment - Overview and application, #5032</i>
	Direct current This symbol indicates on the rating plate that the equipment is suitable for direct current only; to identify relevant terminals. <i>IEC 60417: Graphical symbols for use on equipment - Overview and application, #5031</i>
	Date of Manufacture This symbol indicates the date when the medical device was manufactured. <i>ISO 15223-1. Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General Requirements. #5.1.3</i>
	Manufacturer This symbol indicates the medical device manufacturer. <i>ISO 15223-1. Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General Requirements. #5.1.1</i> Supplemental Product-Specific Manufacturer Information This symbol indicates who the legal manufacturer of the product is.
	Batch Code Indicates product lot number.

Table 1 VT Symbols Glossary (Continued)


Symbol	Description
	Do Not Reuse Indicates product is single use only.
	Use By Date Indicates product expiration date.
	Contents Indicates product contents.
	Catalogue Number This symbol indicates the manufacturer's catalogue number so that the medical device can be identified. <i>ISO 15223-1. Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General Requirements. #5.1.6</i>
	Serial number This symbol indicates the manufacturer's serial number so that a specific medical device can be identified. <i>ISO 15223-1. Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General Requirements. #5.1.7</i>
	Country of Origin Symbol This symbol indicates the country that the product was manufactured in.
	Polarity of d.c. power connector This symbol indicates the positive and negative connections (the polarity) of a direct current power supply, or the positive and negative connections on a piece of equipment to which a direct current power supply may be connected. <i>IEC 60417: Graphical symbols for use on equipment - Overview and application, #5926</i>
Info for USA only: California Proposition 65 	California Proposition 65 This symbol indicates that this product can expose you to chemicals known to the State of California to cause Cancer and Reproductive Harm. For more information go to https://www.P65Warnings.ca.gov.

Table 1 VT Symbols Glossary (Continued)

Symbol	Description
	Temperature limitation Indicates storage requirements limit.
	Start button The Start button initiates a pre-timed cycle at a fixed speed.  Note The StatSpin VT has no on-off switch, and therefore is normally left plugged in and “on”.
	Stop or Open button The Stop or Open button interrupts the cycle and stops the centrifugation. This button can also be used to release the cover.
	Cycle Selector button The Cycle Selector button allows for selection of the cycle.
	Error or Service Indicator The red LED, identified as Error or Service, is illuminated continuously or flashing when service is required.

Contents

	Revision History 55-003786-001, iii
	Safety Notice, v
CHAPTER 1:	Installation, 1-1
	Inspecting the Packaging, 1-1
	Confirming the Contents, 1-1
	Installing the System, 1-1
	Connecting the Power, 1-2
CHAPTER 2:	Overview, 2-1
	Intended Use, 2-1
	Product Description, 2-1
	Operating Controls, 2-1
	Error Indicators, 2-2
	Accessories, 2-3
CHAPTER 3:	Operating Instructions, 3-1
	Opening and Closing the Cover, 3-1
	Installing the Rotor, 3-1
	Description of Rotors, 3-3
	RT12 Tube Rotor, 3-3
	Installing a Shield, 3-3
	RH12 Microhematocrit Rotor, 3-3
	Cycle Selection, 3-4
	Instructions for Use, 3-4
CHAPTER 4:	Specimen Processing , 4-1
	Preparation of Urine Sediment for Microscopic Examination, 4-1
	Determination of Packed Red Cell Volume or Microhematocrit, 4-1
	Procedure for Packed Red Cell Volume (Microhematocrit), 4-2
	Quality Control, 4-3
	Normal Values, 4-3
	Illustrations, 4-3
CHAPTER 5:	Maintenance, 5-1
	Overview, 5-1
	Cleaning, 5-1
	Inspecting the Rotor Speed, 5-2
	Replacing the O-Ring, 5-2
	Troubleshooting, 5-2
	Service, 5-3

Contents

APPENDIX A:	Specifications , A-1
	Specifications , A-1
APPENDIX B:	References , B-1
	References , B-1

Inspecting the Packaging

The StatSpin VT and its accessories are delivered in one carton. If the centrifuge or accessories have suffered any damage in transport, inform your carrier immediately.



Note

Save shipping carton and components to simplify return if service is required.

Confirming the Contents

Product No.:

- SSVT (no rotors supplied)
- SSVT-1 (supplied with two rotors, RT12 and RH12)

Each of the packages contains:

- One External Power Supply (Beckman Coulter Part No. X01-003553-001)
- One grounded line cord (for North American use only)
- One Instructions for Use
- One Sample Pack of various consumable products
- One Warranty Registration Card – Complete the warranty registration as directed.

Installing the System

To install the system:

- 1 Place the StatSpin VT on a level surface suitable for operation of laboratory instrumentation.



Caution

During operation maintain a 30 cm (12 inch) clearance around the centrifuge. The clearance must be free from obstruction and away from the edge of the surface that the centrifuge is on. The appliance coupler on the external power supply is considered the disconnect device, ensure that it is accessible after installation.

- 2 Position the StatSpin VT away from direct sunlight and sources of heat or cold. For the acceptable range of operating temperature and humidity, refer to [Specifications](#).

Installation

Connecting the Power

Connecting the Power

Plug the power cord of the external power supply into a grounded outlet supplying the voltage and frequency indicated on the power supply. When power is connected, the Quick Spin setting LED illuminates, two beeps sound, and the cover lock releases. To turn the system off completely, disconnect the power located at the rear of the unit.



Warning

Only use the external power supply (Beckman Coulter Part Number X01-003553-001) included with the unit. Use of any other external power supply can cause damage to the unit and will void the warranty.



Warning

Outside of North America: do not use the power cord supplied. Use power cord for at least 1.0 Amp with an IEC320/CEE22 female connector and male connector suitable for the power outlet to be used.

Intended Use

To produce the rapid separation of whole blood, preparing urine sediment for microscopic analysis and centrifuging microhematocrit tubes for packed cell volume determination.

Product Description

The StatSpin VT is a small, quiet, high speed centrifuge. Its unique, proprietary drive and suspension system provides nearly vibration-free operation. Light-weight, low mass rotors achieve both top speed and full braking in a few seconds. This instrument is designed to meet international safety standards.

Operating Controls

Table 2 VT Operator Controls

	Start button The Start button initiates a pre-timed cycle at a fixed speed.  Note The StatSpin VT has no on-off switch, and therefore is normally left plugged in and “on”.
	Stop or Open button The Stop or Open button interrupts the cycle and stops the centrifugation. This button can also be used to release the cover.
	Cycle Selector button The Cycle Selector button allows for selection of the cycle.
	Error or Service Indicator The red LED, identified as Error or Service, is illuminated continuously or flashing when service is required.

 Note


Buttons should be depressed with finger tips only. Never press buttons with a sharp object such as a pen, screwdriver, centrifuge insert, fingernail, etc. The buttons are membrane switches designed to be activated by finger actuation. Use with any hard, sharp object can cause damage to the tactile layer of the button, rendering the button unstable and prone to premature failure.

Error Indicators

StatSpin VT error conditions are signaled by a flashing red Error/Service indicator on the control panel.

The combination of the Error or Service indicator and cycle LEDs on the front panel specify the error code.

Figure 1 Rotor failed to reach full speed in 30 seconds



1. Error or Service indicator flashing or continuous

Figure 2 Cover opened or cycle changed during operation





Figure 3 Unable to maintain rotor speed



The system beeps continuously if one of the following conditions are present:

- Centrifuge is over operating temperature
- Short circuit of the motor drives, fan, or solenoid
- Reduction in the availability of electrical power
- Short-term power failures

To stop the beeps, press the Stop button.

Error or Service indicator continuously illuminates when the centrifuge has achieved a total run time of 150 hours, which is the useful life of the drive system. Drive mechanism needs replacing. Contact Beckman Coulter Customer Support.

Accessories

Table 3 VT Accessories

Product No.	Description	Cycle Required
RT12-KIT	2 x 1.5 mL Fixed Angle Rotor with shield	Quick, Normal, Hard
TU15-10	1.5 mL Pre-calibrated Urine Tube (10 bags of 50)	Urine
RH12	12 Position microhematocrit rotor with circular reader (HR4C)	Microhematocrit
HP8H-10	SafeCrit Capillary Tube (40 mm, Sodium Heparin), 100% plastic microhematocrit tubes for the RH12 rotor. (10 vials of 100)	Microhematocrit
HP8U-10	SafeCrit Capillary Tube (40 mm, Untreated), 100% plastic microhematocrit tubes for the RH12 rotor. (10 vials of 100)	Microhematocrit
HT9H-10	Glass Capillary tube (40 mm Ammonium Heparin), Glass microhematocrit tubes for RH12 rotor. (10 vials of 100)	Microhematocrit

Overview
Accessories

Table 3 VT Accessories (Continued)

Product No.	Description	Cycle Required
HT9U-10	Glass Capillary tube (40 mm Untreated), Glass microhematocrit tubes for RH12 rotor. (10 vials of 100)	Microhematocrit
HS24-10	Sealant Pad for Capillary tubes (10 pads)	N/A
HR05	Hematocrit Reader, Card style for 40 mm hematocrit tubes	N/A
HR4C	Hematocrit Reader, Circular designed to be used in conjunction with the RH12 rotor	N/A
00-Ring	Replacement O-Rings for rotor holder (5 bags of 3)	N/A

CHAPTER 3

Operating Instructions

Opening and Closing the Cover

To close the cover, firmly apply pressure on the cover directly above the latch until the gasket compresses completely and the latch engages with the cover. The electrically operated cover interlock mechanism prevents operation until the cover is completely closed and locked, and prevents the cover from being opened while the centrifuge is in operation. When the cover is completely closed and locked, an operating cycle can be initiated.

The centrifuge has a manually operated latch that holds the cover down after spinning is complete. The interlock is automatically released at the end of the operating cycle or by pushing the **Stop** or **Open** button. Squeeze the black latch pieces together to open cover.

Cover Interlock By-pass



Caution

The cover interlock bypass is for emergency use only. Disconnect the power cord of the external power supply from the electrical outlet and ensure the rotor has come to a complete stop before using the interlock bypass. If the equipment is not used correctly, safety can be impaired.

In case of power failure or malfunction, the electronically operated cover interlock mechanism can be released manually by inserting the straightened end of a large paper clip or similar object into the small hole in the center of the front membrane panel. Manually push the lock lever inward about one inch (25 mm) to release the interlock mechanism if the **Stop** or **Open** button does not release the cover.

Installing the Rotor


- 1 Remove the rotor from accessory box.
- 2 Install rotor by pressing rotor firmly in a downward motion onto the rotor-holder of the StatSpin VT. The rotor bottom fits over a rubber O-Ring on the rotor-holder. The figure shows a rotor (cross-section) in position on the rotor-holder. As the rotor turns, the O-

Operating Instructions

Installing the Rotor

Ring moves outward by centrifugal force, enhancing the frictional coupling between the rotor-holder and the rotor.

Figure 4 Install the Rotor



- 1. Rotor
- 2. Rotor-holder

- 3. O-Ring



Position the RT12 rotor in the centrifuge first, then install the shield.



Do not leave any rotor on the rotor-holder when the centrifuge is not in use for an extended period of time. Doing so may compress the O-Ring and decrease its ability to hold rotors.



Never operate the centrifuge without the rotor properly mounted. Failure to install and secure the rotor correctly can damage the centrifuge.



If rotor is left in position between runs, be certain to completely seat the bottom of the rotor on the holder before spinning another sample. Failure to correctly seat the rotor each time can cause the rotor to become loose during centrifugation.



The RT12 rotor must be balanced before operation. If only one sample is being processed, a second sample can serve as the balance tube, or use a similar tube filled with water as a balance.

Description of Rotors

RT12 Tube Rotor

A 2-place rotor designed to accommodate a variety of sample collection tubes for blood and urine separation, in addition to a variety of standard 1.5 mL and 2.0 mL centrifuge tubes with a maximum diameter of 10.9 mm. These tubes should be supported by the collar of the rotor.



Caution

Do not use glass tubes of any kind in the RT12 rotor.



Caution

Do not operate the centrifuge without a shield in place when using an RT12 rotor.

The following tubes are approved for use with the RT12 rotor:

- Prepared Microtubes: TP1H, TP1U, TP5G and CH03
- Precalibrated Urine Tube: TU15-10
- Standard 1.5 — 2.0 mL microcentrifuge tube (e.g. Eppendorf)

Installing a Shield

A shield should always be used with the RT12 rotor and should be replaced when the rotor is replaced.

Push down on the shield until it is fully seated.



RH12 Microhematocrit Rotor

A 12-place covered rotor used to centrifuge capillary tubes for microhematocrit testing. Maximum tube size is 1.7 mm O.D. x 42 mm L. It is recommended that rubber cushions provided with the rotor be replaced twice a year or whenever a tube breaks in the rotor.

The following tubes are approved for use in the rotor:

Operating Instructions

Cycle Selection

- Glass capillary tubes: HT9H, HT9U
- SafeCrit plastic capillary tubes: HP8H, HP8U



Do not use the shield with the RH12 Microhematocrit rotor.

Consult the insert sheets provided with these rotors for instructions for use. To purchase additional rotors not supplied with your StatSpin centrifuge, contact your local distributor and increase the versatility of your centrifuge.



The RT12 and RH12 rotors have a finite lifespan that is dependent on usage. The RT12 and the shield should be replaced after approximately 3600 cycles, which is equivalent to 18 months of service running an average of 10 cycles per day. The RH12 rotor should be replaced after approximately 2600 cycles, which is equivalent to 12 months of service running an average of 10 cycles per day. Rotors should be inspected routinely and replaced immediately when any crack or visible wear occurs.

Cycle Selection

Select spin time settings to achieve optimum results for specific applications. The following are general guidelines:

Table 4 StatSpin VT Cycle Settings

Setting	RPM/RCF	Time	Rotor
Quick Spin	15,800/12,000	30 seconds	RT12
Normal	15,800/12,000	90 seconds	RT12
Hard Spin	15,800/12,000	150 seconds	RT12
Urine Sediment	9,800/3,900	45 seconds	RT12
Microhematocrit	15,800/13,700	120 seconds	RH12

Instructions for Use

- 1 Lift cover and install rotor. When using RT12 rotor, confirm that the shield is installed.
- 2 To close the cover, firmly apply pressure on the cover directly above the latch until the gasket compresses completely and the latch engages with the cover.
- 3 Select desired cycle by selecting the **SET** button until the correct LED is illuminated.
- 4 Press the **Start** button.

-
- 5** When the cycle is complete, the rotor decelerates to a complete stop in 10 seconds and the latch interlock automatically unlocks.
-
- 6** Squeeze the black latch pieces together to open cover.
-


Operating Instructions

Instructions for Use

Preparation of Urine Sediment for Microscopic Examination

The StatSpin VT quickly prepares urine sediment for microscopic examination. This preparation is accomplished with the pre-calibrated urine tubes. (Product Number TU15)

Figure 5 Calibrated Urine Tube



1. Fill to here
2. Empty to here

-
- 1 Add fresh urine to a urine tube. Fill to the top mark (representing 1.5 mL).
 - 2 Cap the tube using the attached stopper and centrifuge in the Tube Rotor, RT12.
 - 3 Balance the rotor either with another sample or with a water-filled tube.
 - 4 Select the **Urine Sediment** setting.
 - 5 Press the **Start** button.
 - 6 When the cycle is complete, the cover releases. Remove the tube from the rotor and remove the stopper.
 - 7 Invert the tube to drain fluid to the lower mark. (The surface tension retains 0.1 mL.)
 - 8 Recap the tube and resuspend the sediment at the bottom of the tube by holding the tube with the index finger and thumb and flicking the tube with the opposite hand.
 - 9 After sediment has been re-suspended, apply one drop to a microscope slide, apply a cover slip, and examine the slide. Follow the protocol used in your laboratory for urine microscopic examination.
-

Determination of Packed Red Cell Volume or Microhematocrit

Both glass and plastic micro-capillary tubes are available. Product Number HT9H (glass) and HP8H (plastic) have been pretreated with heparin and should be used for capillary

Specimen Processing

Procedure for Packed Red Cell Volume (Microhematocrit)

blood. Store the tubes in a cool dry place. Product Number HT9U (glass) and HP8U (plastic) are untreated and used for venipuncture (anticoagulated) samples.

For glass tubes only: if the vial of tubes is new, unscrew the top, remove and discard the foam cushion, and then reattach the top. Now a single tube at a time can be shaken from the vial through the small hole in the center of the cover.

Procedure for Packed Red Cell Volume (Microhematocrit)

- 1 Capillary blood - sample is collected directly from the puncture site using heparinized tubes, Product No. HT9H or HP8H.

or

Venous blood - take well-mixed anticoagulated blood from a syringe or a vacuum blood sample tube. Use untreated tubes, PN HT9U or HP8U.

- 2 Hold the micro-capillary tube by the end with the color-coded band. For more information, refer to the [Illustration section](#).
-

- 3 Fill to the color-coded band. Remove from sample and tilt the banded end downward until the blood moves half-way between the band and the end of the tube.
-

- 4 Hold the tube in a horizontal position and push the dry (banded) end of the tube fully into the vertically held sealing compound. Twist and remove.
-

- 5 Using a laboratory tissue wipe off any blood that is forced from the other end.
-

- 6 Put the tube, sealed end towards the outer rim, in any of the twelve positions on the Hematocrit Rotor, RH12. This rotor does not need to be balanced. Screw the cover in position.
-

- 7 Holding the rotor by the black "cover knob", attach the rotor to the rotor-holder.

 Important

Always hold hematocrit rotor by the black knob on the rotor cover, when pressing it firmly in a downward motion onto the rotor-holder and when removing the rotor from the centrifuge. Pressing on the outer edges of the rotor may result in damage to the rotor.

- 8 Select the **Microhematocrit** setting.
-

- 9 Press the **Start** button.
-

- 10 After the rotor stops, remove the rotor. Spun tubes inside the rotor can also be read with the circular reader, HR4C or removed from the rotor and read with the card-style reader, HR05.
-

Quality Control

- 1

Follow the Quality Control procedures established for your laboratory.
- 2

To confirm the adequacy of cell packing, perform these steps daily:

a.

Select one or more microhematocrit tubes (preferably with a hematocrit over 50%).

b.

Centrifuge the microhematocrit tube and read.

c.

Spin these microhematocrit tubes a second time.

Confirm that the difference between the initial reading and the second reading is 1% or less.

Normal Values

The following table represents commonly accepted hematocrit values:

Table 5 Animals

Species	%
Canine	37 - 55
Feline	24 - 45
Equine	32 - 52
Bovine	32 - 38
Porcine	32 - 50
Ovine	24 - 45

Illustrations

Figure 6 Filling Capillary Microhematocrit Tube from Vacuum Tube






Figure 7 The Filled Capillary Tube




1. Color-coded band
2. Tilt until blood is positioned as shown

Figure 8 Sealing



1. Push
 - hold tube in a horizontal position
 - hold sealant in a vertical position
 - push tube until bottomed - twist and remove

Figure 9 Position of Blood Sample



1. After Filling

2. After Positioning

3. After Sealing

4. After Spinning
5. Sealant

6. Top of plasma

7. Top of red cells

8. Bottom of red cells

Overview

Beckman Coulter recommends that instrument operators perform periodic inspections and preventative maintenance on all devices. Contact Beckman Coulter at any time if the instrument is not functioning correctly.

 Caution

Disconnect the power cord of the external power supply from the electrical outlet before performing maintenance or inspection.

 Warning

Do not expose the rotor to strong or concentrated acids, bases, esters, aromatic or halogenated hydrocarbons, ketones, or strong oxidizing agents, or environmental influences, including natural ultra-violet radiation. Doing so will subject the rotor to corrosion or weakening of the construction.

Cleaning

Clean the outside surfaces and the control panel with a water-dampened cloth and mild detergent. Clean the inner surface or bowl, with a mild detergent, and if necessary, a disinfectant, wiping with an cloth dampened with 70% alcohol or 10% bleach solution.

 Caution

Do not spray cleaning solutions directly onto the centrifuge bowl or housing. Overspray can reach the motor bearings or internal circuitry, causing harm to the electronics and could also cause corrosion or weakening of the construction of the protective casing.

Before using any cleaning or decontamination methods other than those recommended by the manufacturer, users should check with the manufacturer that the proposed method will not damage the equipment. Cleaning and decontamination

may be necessary as a safeguard before laboratory centrifuges, rotors, and accessories are maintained, repaired, or transferred.

Inspecting the Rotor Speed

The rated speeds can be inspected with a stroboscope or photoelectric tachometer. If the StatSpin VT fails to achieve operating speed ($\pm 5\%$) contact Beckman Coulter Customer Support.


Replacing the O-Ring

The figure illustrates the position of the rubber O-Ring which is attached to the rotor-holder. If the rotor becomes difficult to install, apply a very small amount of silicone-type lubricant to the 3 points on the O-Ring gasket. The O-ring should be inspected regularly and replaced when it appears “flattened” or worn. The O-Ring should be replaced as preventative maintenance at least once a year.

A new O-Ring can be installed as shown, by weaving it behind and in front of the 6 pins on the rotor-holder.

The points at which the O-Ring touches the rotor are indicated by the number 1 in the figure. Extra O-Rings have been included.

Figure 10 Position of the O-Ring



1. Positions where the O-Ring touches the rotor.

Troubleshooting

No Power Light		
Is the instrument completely plugged into the electrical outlet?	No →	Plug power cord of the external power supply into the electrical outlet.
Yes ↓		
Is the electrical outlet working correctly?	No →	Try another electrical outlet.

No Power Light	
Yes ↓	
Is the barrel connector from the power adaptor completely plugged into the centrifuge?	No → Plug the barrel connector of the external power supply into the centrifuge.
Yes ↓	
Contact Beckman Coulter Customer Support.	

Centrifuge Will Not Spin/Shuts Off Prematurely	
Is the cover completely closed and cannot be opened manually?	No → Completely close the cover.
Yes ↓	
Is the error light flashing?	Yes → Identify source of error from flashing indicator lights, correct the error condition, and push the Stop/Open button to clear the error.
No ↓	
Contact Beckman Coulter Customer Support.	

Centrifuge Does not Open at the End of the Cycle
Follow Cover Interlock By-pass instructions in Chapter 3 to open the cover and retrieve samples, then contact Beckman Coulter Customer Support.

Service

Refer all service to qualified service personnel or Contact Beckman Coulter Customer Support at 1-800-854-3633.

Be sure to complete and return the warranty card as directed.

Decontamination before returning for service:

Any instrument or accessory containing accumulated blood or other biological or chemical deposits must be cleaned before shipment for service. This decontamination is required by Federal Law (Title 48 and 49 of the Federal Regulations) and according to the Environmental Protection Agency's Regulations for Biohazard Waste Management. Beckman Coulter cannot perform decontamination.

APPENDIX A

Specifications

Specifications

Product Number	SSVT (no rotors supplied)
	SSVT-1 (supplied with two rotors, RT12 and RH12)
Model Number	V901
Cycle and Speeds	Quick Spin: 30 seconds at 15,800 rpm, 12,000 xg
	Normal: 90 seconds at 15,800 rpm, 12,000 xg
	Hard Spin: 150 seconds at 15,800 rpm, 12,000 xg
	Urine: 45 seconds at 9,800 rpm, 3,900 xg
	Hematocrit: 120 seconds at 15,800 rpm, 13,700 xg
Acceleration Time	Approximately 6 seconds
Deceleration Time	Approximately 10 seconds
Electrical	24Vdc, 1.7A Includes switching power supply for 100 to 240 VAC, 50/60 Hz
Dimensions	Diameter: 6.6 in (16.25 cm)
	Height: 6.3 in (13.2 cm)
	Weight: 5.5 lbs (2.5 kg)
Environmental	Indoor use only (IP20)
	Maximum altitude up to 2,000 m
	Maximum relative humidity 80% for temperatures up to 31°C decreasing linearly to 50% relative humidity at 40°C
	Operating temperature: 15°C to 32°C
	Storage temperature: 0°C to 60°C
	Main supply voltage fluctuations not to exceed $\pm 10\%$ of the nominal voltage
	Transient over-voltages according to installation category II
	Pollution degree 2

Specifications
Specifications

References

1. CLSI. *Protection of Laboratory Workers from Occupationally Acquired Infections; Approved Guideline-Second Edition*. CLSI document M29-A2 [ISBN 1-56238-453-8]. CLSI, 940 West Valley Rd, Suite 1400, Wayne, Pennsylvania 19087-1898 USA, 2001.
2. Centers for Disease Control. *Recommendations for prevention of HIV transmission in health-care settings*. MMWR 1987; 36 (suppl no. 2S): 1-18.
3. Centers for Disease Control. *Updated:U.S. Public Health Service Guidelines for the Management of Occupational Exposures to HIV and Recommendations for Postexposure Prophylaxis*. MMWR 2005; 54: 1-29.
4. CLSI. *Procedure for Determining Packed Cell Volume by the Microhematocrit Method. Approved Standard-Third Edition*. CLSI document H7-A3 [ISBN 1-56238-413-9]. CLSI, 940 West Valley Rd, Suite 1400, Wayne Pennsylvania 19087-1898 USA, 2000.
5. CLSI. *Procedures for the Collection of Diagnostic Blood Specimens by Venipuncture; Approved Standard-Fifth Edition*. CLSI document H3-A5 [ISBN 1-56238-515-1]. CLSI, 940 West Valley Rd, Suite 1400, Wayne, Pennsylvania 19087-1898 USA, 2003.
6. CLSI. *Procedures and Devices for the Collection of Diagnostic Capillary Blood Specimens ; Approved Standard-Fifth Edition* CLSI document H4-A5 [ISBN 1-56238- 538-0]. CLSI, 940 West Valley Rd, Suite 1400, Wayne, Pennsylvania 19087-1898 USA, 2004.
7. CLSI. *Collection, Transport, and Processing of Blood Specimens for Testing Plasma- Based Coagulation Assays; Approved Guideline – Fourth Edition*. CLSI document H21- A4 [ISBN 1-56238-521-6]. CLSI, 940 West Valley Rd, Suite 1400, Wayne, Pennsylvania 19087-1898 USA, 2003.

References
References

www.beckmancoulter.com



© 2025 Beckman Coulter, Inc.
All Rights Reserved.