



OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-25.06 revision 1
Project number 3881188
Page 1 of 2

Issuing authority

NMi Certin B.V.
Person responsible: M.Ph.D. Schmidt

Applicant and
Manufacturer

Shinko Denshi Co., Ltd.
1-52-1 Itabashi, Itabashi-ku
Tokyo 173-0004
Japan

Identification of the
certified type

A Non-automatic weighing instrument
Type : CJC series

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class **II**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
6 October 2025

Certification Board

NMi Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



OIML Member State
The Netherlands

Number R76/2006-A-NL1-25.06 revision 1
Project number 3881188
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMI-3881188-01 revision 1 dated 6 October 2025 that includes 31 pages;
- No. NMI-3881188-02 revision 1 dated 6 October 2025 that includes 36 pages;
- No. NMI-3881188-03 revision 1 dated 6 October 2025 that includes 31 pages;
- No. NMI-3881188-04 revision 1 dated 6 October 2025 that includes 18 pages;
- No. NMI-3881188-05 revision 1 dated 6 October 2025 that includes 21 pages;
- No. NMI-3881188-06 revision 1 dated 6 October 2025 that includes 18 pages;
- No. NMI-3881188-07 revision 1 dated 6 October 2025 that includes 21 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	II	
Maximum capacity	$\text{Max} \leq 820 \text{ g}$ $\text{Max} \leq 0.82 \text{ kg}$ $\text{Max} \leq 4100 \text{ ct}$	$\text{Max} \leq 15000 \text{ g}$ $\text{Max} \leq 15 \text{ kg}$ $\text{Max} \leq 75000 \text{ ct}$
Verification scale interval	$e \geq 0,01 \text{ g}$ $e \geq 0,00001 \text{ kg}$ $e \geq 0,05 \text{ ct}$	$e \geq 0,1 \text{ g}$ $e \geq 0,0001 \text{ kg}$ $e \geq 0,5 \text{ ct}$
Actual scale interval	$e = d$	
Weighing range	Single interval	
Maximum number of scale intervals (one weighing range)	$n \leq 82000$	
Tare	$T \leq -\text{Max}$	
Temperature range	$+5^\circ\text{C} / +35^\circ\text{C}$	
Power supply voltage	$100 - 240 \text{ V AC } 50/60 \text{ Hz}$ to $5,95 \text{ V DC}$ (by AC/DC plug-in power supply), Or 6 V DC (by battery)	
Software identification	Checksum: 8969	

Revision History

Revision	Date	Change(s)
0	2025-06-02	Initial issue.
1	2025-10-06	Correction of typos in reports.